

Make sure, that the following items are included:


- radon detector(s) in a sealed bag
- fillable form
- return envelope

If any of these items are missing or the detector has been opened, please contact Radiation and Nuclear Safety Authority on Tuesdays 10am–1pm tel. (09) 7598 8497 (the telephone service is closed from Midsummer to the end of July and in weeks 51-52).



Starting the measurement:

- please start the measurement **as soon as possible** after you have received the detectors
- **remove the detectors from the plastic bag**, but do not open the detectors
- mark the measurement **starting date**, room type and floor on the **form** or through **STUK e-service***



Place the detector:

- in a living space such as living room or bedroom, on a bookshelf or other open surface at breathing height (the position of the detector does not affect the results)
- on the ground floor if you are using just one detector
- on different floors (but not in storage spaces) if you are using two detectors

Do not place the detector:


- near a door, window or an air vent (minimum distance 1,5 m)
- on a mantelpiece, less than 25 cm from a concrete or stone surface or from a lightweight brick wall
- in wet spaces or on the floor (minimum distance 1 m)
- in a location where there is heat or steam



Ending the measurement:

- mark the measurement **ending date** on the form or through STUK e-service*
- send the detectors and the form to STUK using the return envelope **immediately after ending the measurement***
- one return envelope can hold 3-4 detectors with their forms
- If you send the measurement dates through e-service, you don't need to return the form
- **remember to pay for the postal fee of the return envelope**
- the detectors can also be returned to STUK by regular letter or parcel, in which case the recipient of the shipment can be written as "Radiation and Nuclear Safety Authority, Radon laboratory, Jokiniemenkuja 1, 01370 Vantaa" (tel. +358 50 476 3978, email. radonlaboratorio@stuk.fi)
- the results will be sent to the customer in about one month after STUK has received the detectors*

**If the detectors were ordered by someone else, such as landlord, you need to agree with this person on how you will mark the measurement data, return the detectors and receive the results.*



NOTE!

- The recommend length of the measurement is 3 months between 1.9.-31.5. The length of the measurement must be at least 2 months.
- The ventilation system must be working normally. If the building is uninhabited, the temperature must be at least 10 °C.